

SEXUAL HEALTH COLLECTION KIT

A rectal swab sample collection kit used for the collection and transportation of samples for laboratory analysis of sexually transmitted infections by nucleic acid technique (NAT)

This kit contains the materials required for either sample self-collection or collection by a health/social care professional.



For in vitro diagnostic use, sample self-collection









Keep out of reach of children



Do not re-use

Carefully read the

leaflet before use

enclosed instruction



For use by persons aged 18 and over

Sample collection instructions (Steps 1-11)

- Please ensure the kit is within expiry date and read these instructions carefully and completely before attempting to collect the sample.
- If your kit requires online activation, please follow the instructions provided by your healthcare organisation. If you need assistance please contact the healthcare organisation who arranged the test.

Your sample collection kit contents


Please check that the kit contains all of the items outlined below. Do not proceed with sample taking if any items are missing or damaged, contact the healthcare organisation who arranged your test for assistance.











Protective packing wallet



Sample label

Return security label






Return postal envelope

Collection kit outer box


RECTAL SWAB SAMPLE COLLECTION




Clearly complete a sample collection tube label using a ball point pen with:


- Your First name
- Your Surname
- · Your Date of birth
- · Sample type: Rectal























IMPORTANT CHECKLIST Before you return your sample please do the following: Place your completed request form supplied by your doctor/clinic into the postal envelope Make sure the swab specimen transport tube is labelled and is placed inside the protective packing wallet Place the packing wallet into the collection kit outer box Close the collection kit outer box firmly and apply the return security label to seal it Place the collection kit outer box into the return postal envelope Check that you have taken note of your postal tracking number You are now ready to seal the postal envelope. Please store at room temperature until posted. Please post your sample to The Doctors Laboratory as soon as possible (ideally on same day or within 24 hours of sample collection) from ANY Royal Mail post box in the UK. No stamp is required within the UK. If you need assistance please contact The Doctors Laboratory on 020 7307 7373 or email samples@tdlpathology.com. We would welcome feedback on your experience of using this self-collection kit to help us improve our services. To complete a short online survey please scan the code or visit: https://forms.office.com/r/SWWruqwKda **KEY TO SYMBOLS UK Conformity** For in vitro Use by end IVD diagnostic use of year-month Assessed REF EU Conformity Catalogue number Temperature LOT Batch code Caution, consult Do not re-use accompanying Sterilized using STERILE EO documents ethylene oxide Consult instructions Keep away Biological risk from sunlight For use by persons aged 18 and over Biological sample, treat as potentially infectious



REAL Digital International Limited, 2 Queensway, Croydon, Surrey, CRO 4BD, UK Website: www.real-digital.co.uk

Manufacturer

FU Authorised

Representative

Keep dry



Casus Europe BV, Lange Viestraat 2 B, 3511BK Utrecht, The Netherlands.

EC REP



The Doctors Laboratory Limited

Do not use if

Distributor

package is damaged

Distribution is restricted to regions in which the product is registered by the manufacturer.

© The Doctors Laboratory Limited, 2024. Illustrations: © Jag Matharu/Thin Air Productions RDI-IFU-039, TAP5061C/05-02-24/V9. Issue No. 2; Date 05/02/24.

Revision/Date		Change summary
1	02/2023	Initial RDi product for UK market
2	02/2024	RDi IVD product CE registration

If any serious incident occurs in relation to the use of this kit, it should be reported to the manufacturer and the competent authority of the Member State in which the user and/or the patient is established.

Warnings and precautions

- This kit is designed for use by persons aged 18 and over and upon request of a healthcare professional or healthcare organisation.
- Please consult with a healthcare professional for guidance on sample collection processes for adolescent and younger children.
- The kit should not be used by individuals lacking the physical or mental capacity
 to correctly follow the self-collection instructions. If you have problems or feel
 unwell or lightheaded during the collection process, please pause and if necessary,
 consult with your advising healthcare professional.
- The Hologic transport buffer contains no substances which at their given concentration, are considered to be hazardous to health. In the case of contact with eyes, skin or mouth please and rinse thoroughly with clean water, if ingested please drink plenty of water and if any symptoms occur please consult a medical professional. In the case of spillages please wipe with absorbent paper and dispose in normal household waste. For further information please refer to safety datasheets which can be found at www.hologic.com or contact MolecularSupport@hologic.com.
- Samples arriving at the laboratory which show evidence of the below may not be tested:
 - Leaking sample/no liquid present in collection tube
 - . No swab inside the collection tube
- Two or more swabs inside the collection tube
- · Swab placed upside down in the collection tube
- · Different swab type/brand to the one supplied
- General damage
- The sample collection swab has a score line, which after sampling helps the swab break at the correct position when pressure is applied to allow it to fit in the collection tube. If this score line is broken, appears damaged or the swab arrives in two broken parts, do not use and request another kit from your healthcare organisation.
- Use with caution if allergic to nylon fibre and ABS (Acrylonitrile butadiene styrene) material. If any problems arise during the sample collection process please contact your healthcare organisation.
- The accuracy of your results may be compromised if you do not read and follow the instructions in full.
- Samples arriving at the laboratory which show signs of degradation or general damage or arrive after 60 days (Aptima swab) of sample taking may not be tested.

Materials required but not provided

Test request form. This will be provided by your healthcare professional
or healthcare organisation. Please complete the request form with your
details as instructed.

Laboratory Tests

- The tests and procedures undertaken by The Doctors Laboratory Limited are verified and performed in line with supplier product instructions for use and supported by additional validation data for use with self-collection procedures.
- Test results are provided in line with clinically approved results pathways, agreed between The Doctors Laboratory Limited and the patients designated healthcare professional or healthcare organisation.